

16/04/2015

SOP-100.5G

CERTIFICATION OF STERILIZATION

Sterilization Batch No. 22939


Date of Sterilization: 02/04/2015

Customer: Coolsense medical

<u>No. of boxes</u>	<u>Article No. Lot No.</u>	<u>Delivery Note</u>
4		PO 13240

Mediplast's delivery note : 69408

This ETO-sterilized batch, performed according to the ISO-11135:2014 standard, was in compliance with Mediplast's approved cycle specifications. Sterility test of the B.I.s' used for monitoring this cycle showed no growth.

Signature :
 B. Feit ^{ב. פייט}
 Q.A. 

 Quality Assurance Manager